



All India Institute of Medical Sciences, Bhubaneswar  
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

Reference No : J-11025(027)/2013/S&P

Dated : 9<sup>th</sup> OCT 2013


CORRIGENDUM

Tender Ref. : J -11025(027)/2013/S&P  
Last Date : 15/10/2013  
Name of the item : Double Head Binocular Microscope  
Number of Microscopes : 09 (Nine)

Specifications: Amendment

- a. Abbe condenser with objective grid N.A. 1.25 instead of Swing out type plan apochromatic condenser N.A. 0.90.
- b. Double head binocular for two observer with dual colour LED pointer with joy stick control and intensity control instead of double head for two observation for two persons.
- c. Objective should be plan achromatic instead of objective should be flat apochromatic
- d. C Mount / F Mount Adopter instead of C Mount adapter
- e. High resolution digital connection SLR camera with minimum resolution of 12 mega pixel instead of high resolution digital camera with resolution of minimum of 8 Megapixel.
- f. Eye piece 10 x only instead of eye piece 5 x, and 10x.

All other Terms and conditions remains same.

  
Dr. Binod Kumar Patro  
Store & Purchase Officer I/C

Copy to: Dr Manish Goyal for uploading in the website